



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/596,617	02/05/2007	Thomas Bork Hardahl	742111-174	8471

25570 7590 05/27/2010  
ROBERTS MLOTKOWSKI SAFRAN & COLE, P.C.  
Intellectual Property Department  
P.O. Box 10064  
MCLEAN, VA 22102-8064

EXAMINER
----------

DIETRICH, JOSEPH M

ART UNIT	PAPER NUMBER
----------	--------------

3762

NOTIFICATION DATE	DELIVERY MODE
-------------------	---------------

05/27/2010

ELECTRONIC

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

Notice of the Office communication was sent electronically on above-indicated "Notification Date" to the following e-mail address(es):

lgallaugh@rmsclaw.com  
dbeltran@rmsclaw.com  
rescobar@rmsclaw.com

<b>Office Action Summary</b>	<b>Application No.</b> 10/596,617	<b>Applicant(s)</b> HARDAHL ET AL.	
	<b>Examiner</b> Joseph M. Dietrich	<b>Art Unit</b> 3762	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☒ Responsive to communication(s) filed on 04 February 2010.
- 2a) ☐ This action is **FINAL**.                      2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 1-13 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-13 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 19 June 2006 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All    b) ☐ Some \*    c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

### Attachment(s)

- |   |   |
|---|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892)         | 4) <input type="checkbox"/> Interview Summary (PTO-413)           |
| 2) <input type="checkbox"/> Notice of Draftperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____                                      |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)         | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____   | 6) <input type="checkbox"/> Other: _____                          |

## **DETAILED ACTION**

### ***Response to Arguments***

1. Applicant's arguments with respect to claims 1, 11, and 13 have been considered but are moot in view of the new ground(s) of rejection.

### ***Double Patenting***

2. The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. A nonstatutory obviousness-type double patenting rejection is appropriate where the conflicting claims are not identical, but at least one examined application claim is not patentably distinct from the reference claim(s) because the examined application claim is either anticipated by, or would have been obvious over, the reference claim(s). See, e.g., *In re Berg*, 140 F.3d 1428, 46 USPQ2d 1226 (Fed. Cir. 1998); *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) or 1.321(d) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent either is shown to

Art Unit: 3762

be commonly owned with this application, or claims an invention made as a result of activities undertaken within the scope of a joint research agreement.

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

3. Claims 1 – 13 are rejected on the ground of nonstatutory obviousness-type double patenting as being unpatentable over claims 10 and 14 of U.S. Patent No. 7,477,936. Although the conflicting claims are not identical, they are not patentably distinct from each other because both recite analyzing ECG curvature for indicating Long QT syndrome.

***Claim Rejections - 35 USC § 112***

4. The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

***Claim Rejections - 35 USC § 101***

5. 35 U.S.C. 101 reads as follows:

Whoever invents or discovers any new and useful process, machine, manufacture, or composition of matter, or any new and useful improvement thereof, may obtain a patent therefor, subject to the conditions and requirements of this title.

6. Claim 13 provides for the use of a system for analyzing ECG curvature for test of drugs, but, since the claim does not set forth any steps involved in the method/process, it is unclear what method/process applicant is intending to encompass. A claim is indefinite where it merely recites a use without any active, positive steps delimiting how

Art Unit: 3762

this use is actually practiced. The method steps must be actively and positively claimed.

Claim 13 is rejected under 35 U.S.C. 101 because the claimed recitation of a use, without setting forth any steps involved in the process, results in an improper definition of a process, i.e., results in a claim which is not a proper process claim under 35 U.S.C. 101. See for example *Ex parte Dunki*, 153 USPQ 678 (Bd.App. 1967) and *Clinical Products, Ltd. v. Brenner*, 255 F. Supp. 131, 149 USPQ 475 (D.D.C. 1966).

### ***Claim Rejections - 35 USC § 103***

7. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

8. This application currently names joint inventors. In considering patentability of the claims under 35 U.S.C. 103(a), the examiner presumes that the subject matter of the various claims was commonly owned at the time any inventions covered therein were made absent any evidence to the contrary. Applicant is advised of the obligation under 37 CFR 1.56 to point out the inventor and invention dates of each claim that was not commonly owned at the time a later invention was made in order for the examiner to consider the applicability of 35 U.S.C. 103(c) and potential 35 U.S.C. 102(e), (f) or (g) prior art under 35 U.S.C. 103(a).

Art Unit: 3762

9. Claims 1 – 3, 6, 8, and 11 – 13 are rejected under 35 U.S.C. 103(a) as being unpatentable over Sweeney (US 6,684,100) in view of Shusterman (US PG PUB 2002/0143263).

Regarding **claims 1, 2, 6, 8, 11, and 13**, Sweeney discloses a system comprising input means connected to an ECG source, where the different parameters (features) of a received ECG curvature are indicated and/or isolated for indicating possible symptoms which relates to or are indications of certain diseases, where said diseases are known to influence the ECG curvature and that a first number of selected parameters, are combined (set of features) in at least a first mathematical analysis, where the result of the analysis is represented as a point in a coordinate system, comprising at least two axes, where the system compares the actual placement in the coordinate system with a number of reference parameters stored in the system, for indicating symptoms of diseases having influence on the ECG curvature (Fig. 4; (Col. 6, line 34 – Col. 7, line 47); but fails to teach that the groups of parameters include symmetry, flatness, duration, and/or complexity and that the curvature is analyzed to indicate Long QT Syndrome acquired by drug influence. Shusterman teaches it is known to use groups of parameters including symmetry, flatness, duration and/or complexity (e.g. paragraphs 5 and claim 4; measuring the QT interval would give the time interval from Tstart to Tend); and that the curvature is analyzed to indicate Long QT Syndrome (e.g. paragraph 74) and indicate drug induced changes (e.g. paragraph 55). It would have been obvious to one having ordinary skill in the art at the time the invention was made to modify the analysis as taught by Sweeney with the analysis that

Art Unit: 3762

indicates Long QT Syndrome as taught by Shusterman, since such a modification would provide the predictable results of providing the physician with more information to better treat the patient.

Regarding **claims 3 and 12**, Sweeney discloses the analyzing process is repeated in the system for further selected parameters (next set of features in real time) for indicating further possible diseases or symptoms (Col. 6, line 34 – Col. 7, line 47)

### ***Allowable Subject Matter***

10. Claims 4, 5, 7, 9, and 10 are objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

### ***Conclusion***

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Joseph M. Dietrich whose telephone number is (571)270-1895. The examiner can normally be reached on M-F, 8:00 - 5:00 EST.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Carl Layno can be reached on 571-272-4949. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3762

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/J. M. D./  
Examiner, Art Unit 3762

/Scott M. Getzow/  
Primary Examiner, Art Unit 3762